

## Corrigendum 01

**Tender No:** RCB/GTE/05/24-25/FTIR

**Tender Name:** Supply, installation and commissioning of FTIR Spectrophotometer at Regional Centre for Biotechnology Faridabad 121001.

This has reference to tender number RCB/GTE/05/24-25/FTIR for Supply, installation and commissioning of FTIR Spectrophotometer at Regional Centre for Biotechnology, Faridabad. The following specifications pertaining to the above stated tender have been modified.

Sr. No.	Current	Modified
1.	SNR 35000:1 or better (Peak-to-Peak for 1minute scan).	SNR 35000:1 or better (Peak-to-Peak for 1minute scan) @ 4cm-1
2.	Source : Non-migrating hotspot, Long life source covering entire MIR range	Non-migrating hotspot/Ceramic source, Long life source covering entire MIR range

All other specifications and terms and conditions of the aforementioned tender remain unaltered.

**-Sd xxx-**  
**Tender Inviting Authority**