

**Post Pre-bid clarification / Corrigendum**

The points raised by prospective bidders during the pre-bid meeting held on 29-08-2019 and discussed in detail and following decisions were taken: -

Tender Ref.	Item / Equipment	Original point in tender in Technical Specification	Amendment / New addition
RCB/NIT/12 /19-20/ ZFSH	Zebrafish Standalone	<p><b>Sr. No. 8;</b> The system must be equipped with 1 quit 4000 pump and 1 quit 6000 pump.</p> <p><b>Sr. No. 10;</b> Tanks: Self-cleaning flow design tanks, autoclavable, clear polycarbonate tanks, should be compatible with the racks for easy removal and secure positioning in the system. Tank numbers and sizes:                      A. Complete tank sets (Tanks, lids and baffles): Small (1-2L; Quantity: 24).                      B. Complete tank sets (Tanks, lids and baffles): Medium (2.5-4L; Quantity: 16).                      C. Complete tank sets (Tanks, lids and baffles): Large (6-8L; Quantity: 9).</p> <p><b>Sr. No. 12;</b> Each tank should be provided with a continuous tank siphoning system by means of a siphon.</p>	<p>The system must be equipped with 1 quite 4000 L/hr pump and quite 1 6000 L/hr pump.</p> <p>As per the original specification</p> <p>Each tank should be provided with either continuous tank siphoning system by means of a siphon or any other self -cleaning method.</p>
		<p><b>GUARANTEE/WARRANTEE PERIOD:</b>                      THE TENDERERS MUST PROVIDE FULL SYSTEM WARRANTY FOR 3 YEARS FROM THE DATE OF COMPLETION OF THE SATISFACTORY INSTALLATION.</p>	<p>The bidders must quote full system warranty for 3 years from the OEM. Post-warranty comprehensive maintenance contract for another 5 years must be separately quoted which must include all spares, accessories and labour.</p>