

Pre-bid meeting clarification of High Performance Spectral Flow Cytometer Analyzer, held on 18.08.2023 at 11am

Tender spec. Sr. No.	Tender Specification	Point raised by the Company	Approved/Justification
1	Bench-top high end spectral flow-cytometer with at least five lasers system, all lasers should be simultaneous usable (Blue-488±5 nm, Red-635 ±10 nm, Violet-405 ± 5nm, true UV-355±5 nm and Yellow Green- 560±5 nm).	M/s ThermoFisher Scientific; Bench-top high end spectral cum Conventional Cell Sorter Analyzer flow-cytometer with at least five lasers system, all lasers should be simultaneous usable (Blue-488±5 nm, Red-635 ±10 nm, Violet-405 ± 5nm, true UV-355±5 nm and Yellow Green- 560±5 nm).	No deviation
2	The system should have capability of at least 45 or more fluorescence colour detection channels, in addition to Forward Scatter(FSC) & Side Scatter (SSC) and the same must be field upgradable to more detection channels (> 45 colors) in future.	M/s Cytex; The system should have capability of at least 45 or more fluorescence colour detection channels, in addition to Forward Scatter(FSC) & Side Scatter (SSC).	No deviation
4	The System should be able to acquire at least 25,000 events/ second or more with continuously adjustable flow rate.	M/s Cytex; The System should be able to acquire at least 25,000 events/ second or more with adjustable flow rate.	The System should be able to acquire at least 25,000 events/ second or more with adjustable flow rate.
7	The system should be quoted with modules for standard tubes and 96 well plates for sample acquisition.	M/s ThermoFisher Scientific; The system should be quoted with modules for standard tubes and optional 96 well plates/multi tube loader for sample acquisition.	No deviation
11	A high-dimensional data analysis software should be provided with the system with perpetual licensed.	M/s Cytex; A high-dimensional data analysis software should be provided with the system with 5-year License.	A high-dimensional data analysis software <u>with dongle</u> should be provided with the system.

17	Company must have a direct presence in India with strong after-sales technical and service support; Company must ensure that all faults are attended to within 24 hrs.	M/s Cytek; Company/ Authorized distributor must have a direct presence in India with strong after-sales technical and service support; Company must ensure that all faults are attended to within 24 hrs.	Company/ <u>Authorized distributor</u> must have a direct presence in India with strong after-sales technical and service support; Company must ensure that all faults are attended to within 24 hrs.
Additional request	1) Terms of Delivery; Free Delivery at site	M/s Cytek; CIP New Delhi for Foreign Currency Bids & FOR RCB New Delhi for INR currency Bids	CIP New Delhi for Foreign currency. FOR RCB Faridabad for INR currency.
	2) ; EMD Original BG document	M/s Cytek; EMD Original BG document should be sent to RCB and copy of it should be uploaded during bid submission.	Agreed

manu
Bas
Aut
Sp
Pindhu
As
Shubh
Agal