

Refrigerated Centrifuge, Prebid meeting clarification held on 1.9.23 at 3pm

RCB/NIT/06/23-24/RC dated 21/8/23 OIy = 03

S.No.	Tender specifications	Points Raised by Vision representing Thermo Scientific	Revised specifications
2	Speed: 300 - 15,000 RPM or more	Speed: 300-17500 RPM or more	1.5/2ml X 24 – 15,000 RPM or higher 50mlX4 – 15,000 RPM or higher
3	Capacity: 4X50ml tubes at a time or 1.5/2ml X 24 at a time	Capacity 6X50ml tubes with RPM 9500/12,000g with biocontainment Lid (third party certified at a time or 1.5/2mlX24 with RPM 17500/30200g with Biocontainment lid (third party certified). Both rotors should be auto lock rotor exchange without use of any tools	1.5/2ml X 24 – 15,000 RPM or better 50mlX4 – 15,000 RPM (20,000g) or better
16	Bacteriological incubator	Capacity need to be mentioned	40L or better
17	Warranty	Mention if applicable on all components	Warranty shall be applicable on all quoted components including the Indian component.
S.No.	Tender specifications	Points Raised by JKG Biosciences representing Tomy digital Biology	Approval/Justification
2	Speed: 300 - 15,000 RPM or more	Mention individual RPM	1.5/2ml X 24 – 15,000 RPM or higher 50mlX4 – 15,000 RPM or higher
3	Capacity: 4X50ml tubes at a time or 1.5/2ml X 24 at a time	Mention individual RPM	1.5/2ml X 24 – 15,000 RPM or better 50mlX4 – 15,000 RPM (20,000g) or better
15	Bacteriological incubator	Mention capacity	40L or better
3	Rotors	Mention capacity	Rotors for following tubes to be included 1.5ml/2ml 50ml 15ml

*[Handwritten signature]*

*[Handwritten signature]*

*[Handwritten signature]*