

Corrigendum Pre Bid Clarifications

Tender No: RCB/NIT/09/20-21/TSM

Tender Name: Supply Installation and commissioning of Trinocular Stereozoom Microscope

S. No.	Name of the Agency	Bidding Document Reference(s) (Section number/page number)	Content of RFP requiring clarification	Clarification raised by the bidder	RCB Remarks
1	Leica	Technical Specification Point No.1	Optics should be parallel optics zoom system or better.	Optics should be parallel optics zoom system or <u>Greenough design</u> .	Point accepted
2	Leica, Zeiss and LabIndia	Technical Specification Point No..8	Dual nosepiece to accommodate two objectives at a time.	Completely remove this point from the <u>specifications</u>	Point accepted
3	Leica and Zeiss	Technical Specification Point No. 14	Digital Camera: Digital Color CCD/CMOS camera capable of handling bright field, fluorescence images with 5.0 million pixel or more resolution, pixel size 3.0 μ x 3.0 μ quantum efficiency 60% or more. Live display mode, 12.0 frames/sec @ full resolution, max 29 frame per /sec @ ROI mode; USB3.0 port for attaching camera to desktop.	Digital Camera: Digital Color CCD/CMOS camera capable of handling bright field, fluorescence images with 5.0 million pixel or more resolution, pixel size 3.0 μ x 3.0 μ or less. Live display mode, 12.0 frames/sec @ full resolution, max 29 frame per /sec @ ROI mode; USB 3.0 port for attaching camera to desktop.	Point accepted

Further, bidding deadline may be please extended by one week.

(Dr. Jayashree Motiani)

Dr Prasad Athare

Dr. Prashant Pawar